

## **DECLARATION OF CONFORMITY**

According to Annex II Medical Device Directive 93/42 EEC

Type of equipment:	Warm Air Mattress Device for pre, intra and postoperative use
Brand name or trade mark:	Kanmed WarmCloud
Type designation / Model Nos:	OT-600-001
Manufacturer:	Kanmed AB Gårdsfogdevägen 18 B 168 66 BROMMA SWEDEN
Standards:	IEC 60 601-1 (2005): Medical electrical equipment - General requirements for the basic safety and essential performance IEC 80 601-2-35 (2009): Medical electrical equipment - Part 2-35: Particular requirements for the basic safety and essential performance of heating devices using blankets, pads or mattresses and intended for heating in medical use.
Test reports:	TR120829 TR120830
Class	I Ib

We hereby declare that the above mentioned devices comply with the national legalisation of LVFS 2003:11 transposing European Medical Devices Directive 93/42/EEC and the Swedish Medical Devices Act (1993:584)

CE marking introduced 2004

Bromma  
2012-08-30

Authorized person:



Lars Karlsson  
Technical Manager