

## DECLARATION OF CONFORMITY

According to Annex II Medical Device Directive 93/42 EEC

Type of equipment: Patient Warmer for Hospital use.

Brand name or trade mark: Kanmed Operatherm

Type designation / Model Nos: Control Unit, **OP3-022, OP3-115**  
Standard Heating Pad, **OP3-104,**  
Short Heating Pad, **OP3-063**  
Long Heating Pad, **OP3-150**  
Paediatric Heating Pad **OP3-050**

Manufacturer: Kanmed AB  
Gårdsfogdevägen 18 B  
168 66 BROMMA  
SWEDEN

Standards: IEC 60 601-1 (2005): Medical electrical equipment - General requirements for the basic safety and essential performance  
IEC 80 601-2-35 (2009): Medical electrical equipment - Part 2-35: Particular requirements for the basic safety and essential performance of heating devices using blankets, pads or mattresses and intended for heating in medical use.

Test reports: TR120620  
TR120822


MDD: Class IIb

We hereby declare that the above mentioned devices comply with the national legalisation of LVFS 2003:11 transposing European Medical Devices Directive 93/42/EEC and the Swedish Medical Devices Act (1993:584) .

CE marking introduced 2008

Bromma  
2015-07-03

Authorized person:



Lars Karlsson  
Technical Manager